



MASSACHUSETTS

Blue Cross Blue Shield of Massachusetts is an independent licensee of the Blue Cross and Blue Shield Association

Medical Policy

Handheld Radiofrequency Spectroscopy for Intraoperative Assessment of Surgical Margins during Breast-Conserving Surgery

Table of Contents

- [Policy: Commercial](#)
- [Policy: Medicare](#)
- [Authorization Information](#)
- [Coding Information](#)
- [Description](#)
- [Policy History](#)
- [Information Pertaining to All Policies](#)
- [References](#)

Policy Number: 546

BCBSA Reference Number: 7.01.140

NCD/LCD: N/A

Related Policies

None

Policy

Commercial Members: Managed Care (HMO and POS), PPO, and Indemnity Medicare HMO BlueSM and Medicare PPO BlueSM Members

Handheld radiofrequency spectroscopy for intraoperative assessment of surgical margins during breast-conserving surgery is [INVESTIGATIONAL](#).

Prior Authorization Information

Inpatient

- For services described in this policy, precertification/preauthorization **IS REQUIRED** for all products if the procedure is performed **inpatient**.

Outpatient

- For services described in this policy, see below for products where prior authorization **might be required** if the procedure is performed **outpatient**.

	Outpatient
Commercial Managed Care (HMO and POS)	This is not a covered service.
Commercial PPO and Indemnity	This is not a covered service.
Medicare HMO Blue SM	This is not a covered service.
Medicare PPO Blue SM	This is not a covered service.

CPT Codes / HCPCS Codes / ICD Codes

Inclusion or exclusion of a code does not constitute or imply member coverage or provider reimbursement. Please refer to the member's contract benefits in effect at the time of service to determine coverage or non-coverage as it applies to an individual member.

Providers should report all services using the most up-to-date industry-standard procedure, revenue, and diagnosis codes, including modifiers where applicable.

CPT Codes

There is no specific CPT code for this service.

Description

As part of the treatment of localized breast cancer, breast-conserving surgery is optimally achieved by attaining tumor-free margins around the surgical resection site. Failure to achieve clear margins will often require additional surgery to re-excise breast tissue. Currently, histologic examination of excised tissues after completion of surgery is the only method to determine definitively whether clear margins were achieved. Intraoperative methods of assessing surgical margins, such as specimen imaging, frozen section pathology, and touch print cytology, are either not highly accurate, not commonly available, or require considerable time and resources.

A device to detect positive margins should have a high sensitivity, indicating the ability to accurately detect any tumor found in the margins, ideally above 95%. While specificity is less important, excess false positive margin detection would lead to additional unnecessary tissue removal. A new device should have a specificity at least matching current standard best practices, estimated at 85%.¹

The MarginProbe is an intraoperative device which uses radiofrequency spectroscopy to measure the dielectric properties of tissue into which it comes in contact. Cancer cells and normal breast tissues produce different signals. A handheld probe is applied to a small area of the lumpectomy specimen and analyzes whether the tissue is likely malignant or benign. The device gives a positive or negative reading for each touch. If any touch on a particular margin gives a positive reading, the margin is considered to be positive and more tissue should be re-excised if possible. The device can only be used on the main lumpectomy specimen; it cannot be used on shavings or in the lumpectomy cavity of the patient's breast. Use of MarginProbe is intended to increase the probability that the surgeon will achieve clear margins in the initial surgery, thus avoiding the need for a second procedure to excise more breast tissue.

Summary

As part of the treatment of localized breast cancer, breast-conserving surgery is optimally achieved by attaining tumor-free margins around the surgical resection site. Handheld radiofrequency spectroscopy for intraoperative assessment of surgical margins (eg, MarginProbe) is intended to increase the probability that the surgeon will achieve clear margins in the initial procedure, thus avoiding the need for a second surgery to excise more breast tissue.

For individuals who have localized breast cancer or ductal carcinoma in situ undergoing breast-conserving surgery (lumpectomy) who receive handheld radiofrequency spectroscopy for intraoperative assessment of surgical margins (eg, MarginProbe), the evidence includes a randomized trial, several historical control studies, and a systematic review. Relevant outcomes are change in disease status and morbid events. In the randomized trial, histologic examination of surgical margins was not used in the control arm; the outcome measure (complete surgical resection) was not directly clinically relevant and was biased against the control arm; and patient follow-up was insufficient to assess local recurrence rates. The difference in re-excision rates between the 2 trial arms was not statistically significant. Diagnostic characteristics of the device showed only moderate sensitivity and poor specificity; thus, the device will miss some cancers and provide frequent false-positive results. Although several historical control studies have shown lower re-excision rates among patients in whom MarginProbe was used, the studies lacked adequate rigor to demonstrate whether the outcomes are attributable to MarginProbe. The studies did not report recurrence outcomes, which is important for assessing adequacy of resection. A randomized trial that assesses recurrence rates is required to evaluate whether the net health outcome

improves with handheld radiofrequency spectroscopy compared with standard intraoperative surgical margin evaluation, including histologic techniques. The evidence is insufficient to determine the effects of the technology on health outcomes.

Policy History

Date	Action
4/2019	BCBSA National medical policy review. Description, summary and references updated. Policy statements unchanged.
3/2018	New references added from BCBSA National medical policy. Background and summary clarified.
3/2017	New references added from BCBSA National medical policy.
11/2015	New references added from BCBSA National medical policy.
10/2014	New references added from BCBSA National medical policy.
2/2014	New medical policy describing investigational indications. Effective 2/1/2014.

Information Pertaining to All Blue Cross Blue Shield Medical Policies

Click on any of the following terms to access the relevant information:

[Medical Policy Terms of Use](#)

[Managed Care Guidelines](#)

[Indemnity/PPO Guidelines](#)

[Clinical Exception Process](#)

[Medical Technology Assessment Guidelines](#)

References

1. Maloney BW, McClatchy DM, Pogue BW, et al. Review of methods for intraoperative margin detection for breast conserving surgery. *J Biomed Optics*. 2018;23(10). PMID 30369108
2. Blue Cross and Blue Shield Association Technology Evaluation Center (TEC). Handheld radiofrequency spectroscopy for intraoperative margin assessment during breast-conserving surgery. *TEC Assessments*. 2013;Volume 28:Tab 4.
3. Schnabel F, Boolbol SK, Gittleman M, et al. A randomized prospective study of lumpectomy margin assessment with use of MarginProbe in patients with nonpalpable breast malignancies. *Ann Surg Oncol*. May 2014;21(5):1589-1595. PMID 24595800
4. Rivera RJ, Holmes DR, Tafta L. Analysis of the impact of intraoperative margin assessment with adjunctive use of marginprobe versus standard of care on tissue volume removed. *Int J Surg Oncol*. 2012;2012:868623. PMID 23326653
5. U.S. Food and Drug Administration (FDA). Summary of Safety and Effectiveness Data (SSED): MarginProbe System. 2012; https://www.accessdata.fda.gov/cdrh_docs/pdf11/P110014B.pdf. Accessed January 26, 2018.
6. Allweis TM, Kaufman Z, Lelcuk S, et al. A prospective, randomized, controlled, multicenter study of a real-time, intraoperative probe for positive margin detection in breast-conserving surgery. *Am J Surg*. Oct 2008;196(4):483- 489. PMID 18809049
7. Butler-Henderson K, Lee AH, Price RI, et al. Intraoperative assessment of margins in breast conserving therapy: a systematic review. *Breast*. Apr 2014;23(2):112-119. PMID 24468464
8. St John ER, Al-Khudairi R, Ashrafian H, et al. Diagnostic accuracy of intraoperative techniques for margin assessment in breast cancer surgery: a meta-analysis. *Ann Surg*. Feb 2017;265(2):300-310. PMID 27429028
9. Gray RJ, Pockaj BA, Garvey E, Blair S. Intraoperative margin management in breast-conserving surgery: a systematic review of the literature. *Ann Surg Oncol*. 2018;25:18-27. PMID 28058560
10. Thill M, Dittmer C, Baumann K, et al. MarginProbe(R)--final results of the German post-market study in breast conserving surgery of ductal carcinoma in situ. *Breast*. Feb 2014;23(1):94-96. PMID 24291375
11. Thill M, Roder K, Diedrich K, et al. Intraoperative assessment of surgical margins during breast conserving surgery of ductal carcinoma in situ by use of radiofrequency spectroscopy. *Breast*. Dec 2011;20(6):579-580. PMID 21885281

12. Sebastian M, Akbari S, Anglin B, et al. The impact of use of an intraoperative margin assessment device on re- excision rates. Springerplus. Apr 28 2015;4:198. PMID 26020017
13. Blohmer JU, Tanko J, Kueper J, et al. MarginProbe(c) reduces the rate of re-excision following breast conserving surgery for breast cancer. Arch Gynecol Obstet. Aug 2016;294(2):361-367. PMID 26796680
14. Coble J, Reid V. Achieving clear margins. Directed shaving using MarginProbe, as compared to a full cavity shave approach. Am J Surg. Apr 2017;213(4):627-630. PMID 28049561
15. Kupstas A, Ibrar W, Hayward RD, et al. A novel modality for intraoperative margin assessment and its impact on re-excision rates in breast conserving surgery. Am J Surg. Nov 21 2017. PMID 29191356
16. National Comprehensive Cancer Network (NCCN). NCCN Clinical practice guidelines in oncology: breast cancer. Version 3.2017. https://www.nccn.org/professionals/physician_gls/pdf/breast.pdf. Accessed January 24, 2018. Conserving Surgery
17. American Society of Breast Surgeons. Performance and Practice Guidelines for Breast-Conserving Surgery/Partial Mastectomy. https://www.breastsurgeons.org/statements/guidelines/PerformancePracticeGuidelines_Breast-ConservingSurgery-PartialMastectomy.pdf. Accessed January 9, 2019.